

File Name: Declaration of conformity

Edition: H

File No.: CS/CE-MD300C-01

Total 2 page 1

Declaration of Conformity to Council Directive 93/42/EEC Concerning Medical Devices

Manufacturer: Beijing Choice Electronic Technology Co., LTD.

Room 4104, No.A12 Yuquan Road, Haidian District,

100143 Beijing, P.R.CHINA

European Representative: Shanghai International Holding Corp. GmbH (Europe)

Eiffestraβe 80, 20537 Hamburg Germany

Product Name: Fingertip Pulse Oximeter

Product Model: MD300C15D, MD300CH3

UMDNS Code: 17148

Classification - Annex IX: Class IIa, rule 10 to Annex IX of the MDD

Conformity assessment Route: Annex II excluding (4)

We, <u>the manufacturer</u>, herewith declare that the stated medical devices meet the transposition into national law, the provisions of Council Directive 93/42/EEC concerning medical devices including the amendments by Council Directive 2007/47/EEC.

All supporting documentation is retained at the premises of the manufacturer.

Standards applied:

EN ISO 13485:2012 Medical devices - Quality management systems - Requirements for regulatory purposes

EN ISO14971:2012 Medical devices –Application of risk management to medical devices EN ISO10993-1:2009 Biological evaluation of medical devices-part 1: evaluation and testing

EN ISO10993-5:2009 Biological evaluation of medical devices-Part 5: Test for in vitro cytotoxicity

EN ISO10993-10:2010 Biological evaluation of medical devices-Part 10: Tests for irritation and delayed-type hypersensitivity

EN60601-1:1990_A1:1993_A2:1995 Medical electrical equipment-Part 1: General requirements for safety

EN60601-1-2:2007 Medical electrical equipment-Part 1-2: General requirements for basic safety and essential performance-Collateral standard: Electromagnetic compatibility-Requirements and tests

EN60601-1-4: 1996/A1:1999 Medical electrical equipment –Part 1-4: General requirements for safety-Collateral standard: Programmable electrical medical systems

EN ISO9919:2009 Medical electrical equipment-Particular requirements for the basic safety and essential performance of pulse oximeter equipment for medical use (ISO9919:2005)

EN1041:2008 Information supplied by the manufacture of medical device

EN 980:2008 Symbols for use in the labelling of medical devices

EN 62304: 2006 Medical device software-Software life-cycle processes



File Name: Declaration of conformity

File No.: CS/CE-MD300C-0

Edition: H

Total 2 page 2

EN60601-1-6: 2010 Medical electrical equipment-Part 1-6: General requirements for basic

safety and essential performance-Collateral Standard: Usability

Notified Body: TÜV SÜD Product service GmbH

Ridlerstr 65, D-80339 München, Germany

Identification number **C€** 0123

(EC) Certificate(s): No. G1 16 02 78179 025

Start of CE-marking: 2006-06-01

Place, Date of Declaration: Beijing, 2016-02-19

Signature:

Name: Lei Chen

Position: Quality Director

Lez Chen