Declaration of Conformity to Council Directive 93/42/EEC concerning Medical Devices

Manufacturer:	Beijing Choice Electronic Technology Co., Ltd. 2nd Floor, 3rd Floor and Room 410-412 4th Floor, No. 2 Building, No. 9 Shuangyuan Road,Shijingshan District, 100041 Beijing, PEOPLE'S REPUBLIC OF CHINA
European Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestraβe 80 20537 Hamburg GERMANY
Product Name：	Fingertip Pulse Oximeter
Product Model:	See attached list
UMDNS Code：	17148
Classification：	Class IIa, rule 10 to Annex IX of the MDD
Conformity assessment Route：	Annex II excluding (4)

We, the manufacturer, herewith declare that the stated medical devices meet the transposition into national law, the provisions of Council Directive 93/42/EEC concerning medical devices.
All supporting documentation is retained at the premises of the manufacturer.

Standards applied:

EN ISO 13485:2016/AC:2016 Medical devices- Quality management systems- Requirements for regulatory purposes

EN ISO14971:2012 Medical devices －Application of risk management to medical devices

EN 60601-1:2006/A1:2013 Medical electrical equipment-Part 1: General requirements for safety

EN 60601-1-2:2015 Medical electrical equipment Part 1-2: General requirements for basic safety and essential performance - Collateral Standard：Electromagnetic disturbances - Requirements and tests

EN 60601-1-6:2010 Medical electrical equipment -- Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability

EN 60601-1-11:2010 Medical electrical equipment -- Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.

ISO 80601-2-61:2011 Medical electrical equipment —Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment

EN ISO10993-1:2009/AC:2010 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process

EN ISO10993-5:2009 Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity

ISO10993-10:2010 Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization

EN1041:2008 Information supplied by the manufacture of medical devices

EN ISO15223-1:2016 Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements

EN 62304:2006/AC:2008 Medical device software-Software life-cycle processes

MEDDEV 2.7/1: 2016 CLINICAL EVALUATION: A GUIDE FOR MANUFACTURERS AND NOTIFIED BODIES UNDER DIRECTIVES 93/42/EEC and 90/385/EEC

Notified Body:	TÜV SÜD Product service GmbH Ridlerstr 65, D-80339 München, Germany
Identification Number：	CE 0123
(EC) Certificate(s):	No. G1 057571 0003 Rev.00
Start of CE-marking:	2016-05-06
Place, Date of Declaration:	Beijing, 2020-03-27
Signature:	 Name: Haiying Zhao Position: Quality Director

Attached list

MD300CN310, MD300CN330, MD300CN340, MD300CN350, MD300CN356, MD300CN360, MD300CN130, MD300CN150, MD300CN160
MD300C1, MD300C11, MD300C12, MD300C13, MD300C15, MD300C16, MD300C17, MD300C18, MD300C19, MD300C1B, MD300C1C, MD300C1D, MD300C1E, MD300C1F, MD300C15D
MD300C2, MD300C20, MD300C201, MD300C203, MD300C204, MD300C21, MD300C21C, MD300C22, MD300C221, MD300C23, MD300C25, MD300C26, MD300C29, MD300C2A, MD300C2B, MD300C2D, MD300C2E, MD300C2F
MD300C4, MD300C41
MD300C5, MD300C52, MD300C53, MD300C54
MD300C63, MD300C634, MD300CF3, MD300CH3
LTD800, LTD805
PM A10, PX-100-EU, LTD800, LTD805, ME5, FL313, FL323, TS 370PO, TS 380PO, PM A19, PO 45, MD300CN140, MD300C55