

**DECLARATION OF CONFORMITY  
TO COUNCIL DIRECTIVE 93/42/EEC  
CONCERNING MEDICAL DEVICES**

**MANUFACTURER:** Edan Instruments, Inc.  
#15 Jinhui Road, Jinsha Community, Kengzi Sub-District,  
Pingshan District, 518122 Shenzhen, P.R.China

**EUROPEAN REPRESENTATIVE:** Shanghai International Holding Corp. GmbH  
Eiffestrasse 80 20537 Hamburg Germany

**PRODUCT/MODEL:** **Pulse Oximeter/ H100B**  
*The accessories are used together with the product*

**GMDN [NAME/CODE]:** Pulse oximeter, battery-powered /45607

**CLASSIFICATION:** Class II b, Rule 10 According To Annex IX OF the MDD

**CONFORMITY ASSESSMENT ROUTE:** Annex II excluding (4)

WE, EDAN INSTRUMENTS, INC., HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT(S) MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 INCLUDING AMENDMENTS BY DIRECTIVE 2007/47/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER.

**STANDARDS APPLIED:** EN 60601-1:2006+A1:2013, EN 60601-1-2:2015, EN 60601-1-6:2010+A1: 2015, EN ISO 80601-2-61:2011, EN ISO 10993-1:2009, EN ISO 10993-5:2009, EN ISO 10993-10:2013, EN ISO 14155:2011, EN ISO 14971:2012, EN 62304: 2006+A1: 2015, EN 62366-1: 2015, EN 15223-1: 2016, EN 1041: 2008+A1:2013, EN ISO 780:2015


**NOTIFIED BODY:** TÜV SÜD PRODUCT SERVICE GMBH  
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

**IDENTIFICATION NUMBER** 0123

**(EC) CERTIFICATE(S):** G1 091264 0006 Rev. 01 VALID UNTIL: 2022-09-17

**START OF CE-MARKING:** 2008-01-04

**PLACE, DATE OF ISSUE:** SHENZHEN, 2019.9.9

**SIGNATURE:**   
NAME LIU YONGYING  
MANAGEMENT REPRESENTATIVE



## DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

**MANUFACTURER:** Edan Instruments, Inc.  
#15 Jinhui Road, Jinsha Community, Kengzi Sub-District,  
Pingshan District, 518122 Shenzhen, P.R.China

**EUROPEAN REPRESENTATIVE:** Shanghai International Holding Corp. GmbH  
Eiffestrasse 80 20537 Hamburg Germany

**PRODUCT/MODEL:** **Pulse Oximeter/ H100N**  
*The accessories are used together with the product*

**GMDN [NAME/CODE]:** Pulse oximeter, battery-powered /45607

**CLASSIFICATION:** Class II b, Rule 10 According To Annex IX OF the MDD

**CONFORMITY ASSESSMENT ROUTE:** Annex II excluding (4)

WE, EDAN INSTRUMENTS, INC., HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT(S) MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 INCLUDING AMENDMENTS BY DIRECTIVE 2007/47/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER.

**STANDARDS APPLIED:** EN 60601-1:2006+A1:2013, EN 60601-1-2:2015, EN 60601-1-6:2010+A1: 2015, EN ISO 80601-2-61:2011, EN ISO 80601-2-56:2012, EN ISO 10993-1:2009, EN ISO 10993-5:2009, EN ISO 10993-10:2013, EN ISO 14155:2011, EN ISO 14971:2012, EN 62304: 2006+A1: 2015, EN 62366-1: 2015, EN 15223-1: 2016, EN 1041: 2008+A1:2013, EN ISO 780:2015


**NOTIFIED BODY:** TÜV SÜD PRODUCT SERVICE GMBH  
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

**IDENTIFICATION NUMBER** 0123

**(EC) CERTIFICATE(S):** G1 091264 0006 Rev. 01 VALID UNTIL: 2022-09-17

**START OF CE-MARKING:** 2008-04-30

**PLACE, DATE OF ISSUE:** SHENZHEN, 2019.9.9.

**SIGNATURE:**   
NAME LIU YONGYING  
MANAGEMENT REPRESENTATIVE