

**DECLARATION OF CONFORMITY  
TO COUNCIL DIRECTIVE 93/42/EEC  
CONCERNING MEDICAL DEVICES**

MANUFACTURER: Edan Instruments, Inc.  
#15 Jinhui Road, Jinsha Community, Kengzi Sub-District,  
Pingshan District, 518122 Shenzhen, P.R.China

EUROPEAN REPRESENTATIVE: Shanghai International Holding Corp. GmbH (Europe)  
Eiffestrasse 80 D-20537 Hamburg Germany

PRODUCT/MODEL: **PC ECG / SE-1515**  
*The accessories are used together with the product*

GMDN [NAME/CODE]: Interpretive multichannel electrocardiograph / 16231

CLASSIFICATION: Class II a, Rule 10 According To Annex IX of the MDD

CONFORMITY ASSESSMENT ROUTE: Annex II excluding (4)

WE, EDAN INSTRUMENTS, INC., HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT(S) MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 INCLUDING AMENDMENTS BY DERECTIVE 2007/47/EC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER.

STANDARDS APPLIED: **EN 60601-1:2006+A1:2013, EN 60601-1-2: 2015, EN 60601-1-6:2010+A1:2015, EN 60601-2-25:2015, EN ISO 10993-1:2009, EN ISO 10993-5:2009, EN ISO 10993-10:2013, EN ISO 14155:2011, EN ISO 14971:2012, EN 62304:2006+A1:2015, EN 62366-1:2015, EN ISO 15223-1:2016, EN 1041:2008+A1:2013, EN ISO 780: 2015**


NOTIFIED BODY: TÜV SÜD PRODUCT SERVICE GMBH  
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER 

(EC) CERTIFICATE(S): G1 091264 0006 Rev. 01 VALID UNTIL: 2022-09-17

START OF CE-MARKING: 2014-03-04

PLACE, DATE OF ISSUE: SHENZHEN, 2019.2.26.

SIGNATURE:   
NAME LIU YONGYING  
MANAGEMENT REPRESENTATIVE