

SCHILLER MEDICAL SAS
ZAE SUD
BP 90050
4, rue Louis Pasteur
F-67162 Wissembourg Cedex

Téléphone +33 (0) 3 88 63 36 00
Télécopie +33 (0) 3 88 94 12 82

SCHILLER
MEDICAL S.A.S.

Internet : <http://www.schiller-medical.com>
E.mail : info@schiller.fr

Wissembourg, 16.01.2015

RoHS II Declaration of Conformity

Referring to the Directive 2011/65/EU on the restriction of the use of certain hazardous substances (RoHS) in Medical Electrical and Electronic Equipment, we declare the conformity of the :

4-07-0001 EASY BAT

The status of compliance with the RoHS II directive is specified; this status is defined compliant, as it doesn't contain the restricted substances above the official maximum concentration value.

Restricted substances and maximum concentration values tolerated by weight in homogeneous materials:	
Lead (Pb)	0,1% by weight in homogeneous materials *
Mercury (Hg)	0,1% by weight in homogeneous materials
Cadmium (Cd)	0,01% by weight in homogeneous materials
Hexavalent chromium (Cr(VI))	0,1% by weight in homogeneous materials
Polybrominated biphenyls (PBB)	0,1% by weight in homogeneous materials
Polybrominated diphenyl ethers (PBDE)	0,1% by weight in homogeneous materials

*'Homogeneous material' means one material of uniform composition throughout or a material, consisting of a combination of materials, that cannot be disjointed or separated into different materials by mechanical actions such as unscrewing, cutting, crushing, grinding and abrasive processes.

Date: 16/01/2015

Name: Alain WEISSINGER

Position: Quality Manager

Signature:

SCHILLER MEDICAL SAS
B.P. 90050
F-67162 WISSEMBOURG Cedex
Tél. (33) 03 88 63 36 00
Fax (33) 03 88 54 36 32



THE ART OF DIAGNOSTICS