SAP DIR: 80019030 Version: B Page 1 of 3



DECLARATION OF CONFORMITY

(in accordance with ISO/IEC 17050-1)

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- the European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices
- the Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonization of the laws of the Member States relating to the making available on the market of radio equipment
- the Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

(RoHS)			
Manufacturer's Name and Business Address:	Welch Allyn, Inc. 4341 State Street Road Skaneateles Falls, NY 13153, USA		
EC REP	Regulatory Affairs Representative Welch Allyn Limited Navan Business Park Dublin Road Navan, County Meath Republic of Ireland		
Product Name ^{1,3} :	Connex® Spot Monitor		
REF 1,3	901058 Vital Signs Monitor Core		
#	The Connex Spot Monitor comes in four series 71,73, 74 and 75 7X (SpO2 option) (Temperature option) – (Power Cord) SpO2 options: W= Nonin; C= Covidien; M= Masimo; X= No SpO2 Temperature options: T= SureTemp Plus; E= Braun Pro 6000; X= No Temperature option. (Power Cord) depends on region.		
Radio equipment ² :	Laird WB45NBT (DIR 60086189)		
Object of the declaration ² :	802.11 a/b/g/n Enterprise Wi-Fi + Bluetooth Communications Subsystem		
Accessories and components ² :	Not Applicable		

¹ applicable to the medical devices directive, 93/42/EEC

² applicable to the radio equipment directive, 2014/53/EU

³ applicable to the RoHS directive, 2011/65/EU

WelchAllyn°

SAP DIR: 80019030 Version: B Page 2 of 3

Medical Device Conformity Assessment Route Annex ¹ :	П		
Medical Device Classification ¹ :	IIa		
Medical Device Classification Rules ¹ :	10		
GMDN Code and Term ¹ :	57960 - Multiple physiological parameter spot-check analysis system, clinical		
UMDNS Code and Term ¹ :	25209- Monitor, Physiologic, Vital Signs		
Notified Body ¹ : (CE 0297)	DQS Medizinprodukte GmbH, August-Schanz-Str.21, 60433 Frankfurt am Main EC-certificate No. 314505 MR2		
Standards Applied	Number	Title	
(Standards are applicable to the medical device directive, unless otherwise indicated):	EN 50581 ³	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances	
	EN/IEC 62304	Medical Device Software – Software Life Cycle Processes	
	EN/ISO 1060-3	Non-Invasive Sphygmomanometers – Part 3. Supplementary Requirements for Electro-Mechanical Blood Pressure Measuring Systems	
	EN/IEC 60601-1	Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance	
	EN/IEC 60601-1-2	Medical electrical equipment - Part 1-2: General Requirements for Safety – Collateral Standard: Electromagnetic Compatibility – Requirements and Tests	
	EN/IEC 62366	Medical devices – Application of Usability Engineering to Medical Devices	
	EN/IEC 60601-1-6	Medical Electrical Equipment – Part 1-6: General Requirements for Safety – Collateral Standard: Usability	
	EN/IEC 60601-1-8	Medical Electrical Equipment – Part 1-8: General Requirements for Safety – Collateral Standard: General Requirements, Tests and Guidance for Alarm Systems in Medical Electrical Equipment and Medical Electrical Systems	

 $^{^{\}rm 1}$ applicable to the medical devices directive, 93/42/EEC $^{\rm 2}$ applicable to the radio equipment directive, 2014/53/EU $^{\rm 3}$ applicable to the RoHS directive, 2011/65/EU

Welch Allyn°

SAP DIR: 80019030 Version: B Page 3 of 3

EN/IEC 60601-2- 49	Medical Electrical Equipment – Part 2-49: Particular Requirements for the Safety of Multifunction Patient Monitoring Equipment
IEC 80601-2-30	Medical electrical equipment - part 2-30: particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers
ISO 80601-2-56	Medical electrical equipment - Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement
ISO 80601-2-61	Medical electrical equipment - part 2-61: particular requirements for basic safety and essential performance of pulse oximeter equipment

Authorised Signatory:

Fiona Butler, Manager Regulatory Affairs

iona Butler

Navan

{EU Authorised Representative}

Place of Issue

applicable to the medical devices directive, 93/42/EEC
 applicable to the radio equipment directive, 2014/53/EU
 applicable to the RoHS directive, 2011/65/EU