

**Welch Allyn®** **DECLARATION OF CONFORMITY**  
(in accordance with ISO/IEC 17050-1)

SAP DIR No.: 80016490

Version: G

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- the European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices
- the Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

Manufacturer's Name and Business Address: Welch Allyn, Inc.  
4341 State Street Road  
Skaneateles Falls, NY 13153, USA

**EC REP** Regulatory Affairs Representative  
Welch Allyn Limited  
Navan Business Park  
Dublin Road  
Navan, County Meath  
Republic of Ireland

Product Name: Retinoscope

**REF** 901024, RETINOSCOPE

**#** 18240, 18245, 18300, 18210-BI, 18342-VC

Medical Device Conformity Assessment Route Annex: VII

Medical Device Classification: I

Medical Device Classification Rules: 12

GMDN Code and Term: 32712 - Retinoscope, battery-powered

UMDNS Code and Term: 17840 - Retinoscope

Standards Applied: EN 50581 Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

EN 60601-1 Medical Electrical equipment – Part 1: General requirements for Safety



# DECLARATION OF CONFORMITY

(in accordance with ISO/IEC 17050-1)

SAP DIR No.: 80016490

Version: G

EN 60601-1-2

Medical electrical equipment – Part 1-2: General requirements for safety – Collateral Standard Safety requirements for medical electrical systems

Authorised Signatory:

Fiona Butler, Manager Regulatory Affairs  
{EU Authorised Representative}

Date

Navan

Place of Issue